



TELL-Seq™ WGS Library Prep User Guide

For genome size from 1 Mb to 5 Gb

This User Guide is to be used with TELL-Seq™ Library Reagent Box 1 V1 which contains newly released Suspension Buffer EZ and TELL-Seq™ Library Reagent Box 2 V1 which includes TELL Bead Plex option.

For Research Use Only. Not for use in diagnostic procedures.

Document # 100005 v9

August 2022

This document is proprietary to Universal Sequencing Technology Corporation and is intended solely for the use of its customers in connection with the use of the products described herein and for no other purposes.

The instructions in this document must be followed precisely by properly trained personnel to ensure the proper and safe use of the TELL-Seq™ kit.

UNIVERSAL SEQUENCING TECHNOLOGY CORPORATION DOES NOT ASSUME ANY LIABILITY OCCURRING AFTER INCORRECT USE OF THE TELL-SEQ™ KIT.

©2022 Universal Sequencing Technology Corporation. All rights reserved.

TELL-Seq™ is a trademark of Universal Sequencing Technology Corporation. All other names, logos, and other trademarks are the property of their respective owners.

Revision History

Doc #CO1K001 Rev. A	October 2019	Initial Release
Doc #CO1K001 Rev. B	December 2019	Added storage instructions
Doc #CO1K001 Rev. C	February 2020	Added storage instructions
Doc #100005 v4	June 2020	Reaction volume changes
Doc #100005 v5	August 2020	Included high throughput kits and library amplification cycle number modification
Doc #100005 v6	November 2020	Updated sequencing primer instruction for NovaSeq v1.5 reagents
Doc #100005 v7	March 2022	Larger reaction volume, double SPRI cleanup option for large size library, Index 2 sequences for sample sheet
Doc #100005 v8	March 2022	Updated handling of 2x PCR Master Mix from “vortex mixing” to “carefully mixing”.
Doc #100005 v9	August 2022	Updated kit with Suspension Buffer EZ for ease-of-handling and improved performance; included TELL Bead Plex option for improved compatibility with non-Illumina sequencing systems; added a new set of 96 Multiplex Primers for high throughput applications

Table of Contents

1. Introduction	5
Genomic DNA Input Recommendations	5
TELL-Seq™ Library Prep Kit Update Notification	6
2. Kit Contents	7
TELL-Seq™ Library Prep Kit, Standard Size	7
Box 1 of 2: TELL-Seq™ Library Reagent Box 1 V1	7
Box 2 of 2: TELL-Seq™ Library Reagent Box 2 V1	7
TELL-Seq™ Library Multiplex Primer (1-8) Kit	8
TELL-Seq™ Library Multiplex Primer (9-16) Kit	8
TELL-Seq™ Library Multiplex Primer (17-24) Kit	8
TELL-Seq™ Library Multiplex Primer C-series	9
TELL-Seq™ Illumina® Sequencing Primer Kit	9
TELL-Seq™ Library Prep Kit, HT24	10
Box 1 of 2: TELL-Seq™ Library Reagent Box 1 V1, HT24	10
Box 2 of 2: TELL-Seq™ Library Reagent Box 2 V1, HT24	10
TELL-Seq™ Illumina® Sequencing Primer Kit, HT	11
3. Consumables and Equipment (not provided)	12
Consumables	12
Equipment	12
4. TELL-Seq™ Library Prep Workflow	13
5. Protocol	14
Barcode DNA	14
Stabilize DNA	17
Tagment DNA	18
Wash Beads	19
Amplify Library	21
Clean Up Library	23
Qualify and Quantify Library	27
6. Appendix	30
Ultralow Input TELL-Seq™ Library Preparation	30
TELL-Seq™ Library Structure and Sequencing Scheme	30

Example of Illumina® Sequencing % Base by Cycle Chart with TELL Bead & T-series Index II	31
Illumina® Sequencing Guide	31
Illumina® Sequencing Read Length Recommendation	32
Sequencing Depth Consideration	32
Library Multiplex Primer Index Sequences (i.e. Index 2 Sequences): T-series (8-base)	33
Library Multiplex Primer Index Sequences (i.e. Index 2 Sequences): C-series (10-base)	34
Usage of TELL-Seq™ Multiplex Primer Plate	36

1. Introduction

This user guide explains how to prepare indexed paired-end TELL-Seq™ whole genome sequencing (WGS) libraries using a TELL-Seq™ Library Prep Kit from a genomic DNA sample with genome size ranges from 1 Mb to 5 Gb for subsequent sequencing on an Illumina® sequencing system.

The TELL-Seq™ Library Prep Kit uses an innovative Transposase Enzyme Linked Long-read Sequencing (TELL-Seq™) technology† to prepare a paired-end library to generate barcode linked reads from an Illumina® sequencing system. Coupled with TELL-Seq™ specific analysis software, the linked reads can be easily used for genome wide variant calling, haplotype phasing, structural variation detection, metagenomic studies and *de novo* sequencing assembly.

A TELL-Seq™ Library Prep kit can generate TELL-Seq™ WGS libraries for genome size ranging from 1 Mb to 5 Gb.

- Use 0.5 ng to 5 ng genomic DNA input for standard input procedure
- Use as little as 0.1 ng DNA for ultralow input applications (Appendix)
- Produce barcode linked reads using an Illumina® sequencing system

A TELL-Seq™ Library Prep kit (Box 1 and Box 2) along with TELL-Seq™ Library Multiplex Primer Kits can generate different number of TELL-Seq™ Libraries based on the kit format and sample genome sizes.

Type of Kit	Genome Type (Size)	No. of TELL-Seq™ Libraries
Standard	Small (1Mb to 200Mb)	12
	Medium (200Mb to 1Gb)	6
	Large (1Gb to 5Gb)	4
HT24	Small (1Mb to 200Mb)	72
	Medium (200Mb to 1Gb)	36
	Large (1Gb to 5Gb)	24

Genomic DNA Input Recommendations

Genomic DNA inputs range from 0.1 ng to 5 ng based on the genome size.

High molecular weight (HMW) DNA is critical for successful sequencing.

- For large genomes, minimum sample DNA size should be greater than 40Kb.
- For small genomes, minimum sample DNA size should be greater than 20Kb.
- HMW DNA ranging from 100Kb to 300Kb are optimal material for best whole genome phasing application.
- Avoid breaking the HMW DNA during handling. Remove low molecular weight DNA (identified as a smear less than 10Kb on a gel) in the sample if they present a significant portion in the DNA sample.

† Patent pending.

Use a fluorometric-based method to quantify input DNA. If you use the Qubit dsDNA BR Assay Kit or HS Kit, use at least 2 μL of each DNA sample for a measurement. Avoid methods that only measure total nucleic acid concentration, such as NanoDrop or other UV absorbance methods.

For accurate measurement of HMW DNA concentration, dilute the concentrated DNA to the working concentration (0.2ng/ μL to 1ng/ μL) in a Tris buffer (pH 7.5-8.0) several hours to a day before the concentration measurement and library preparation.

Genomic DNA should be stored in a Tris buffer with pH ranging from 7.5 - 8.0 or a low TE buffer (10mM Tris-HCl, 0.1 mM EDTA, pH 8.0).

For assessing the purity of a DNA sample, the ratio of absorbance measurement at 260 nm to absorbance at 280 nm can be used. This protocol is optimized for DNA with absorbance ratio values of 1.8–2.0. If there is excessive RNA in the DNA sample, it should be removed with a RNase treatment.


TELL-Seq™ Library Prep Kit Update Notification

We have made further enhancements to TELL-Seq™ library prep kits. TELL-Seq™ Library Reagent Box 1 V1 replaces previous Suspension Buffer with newly released Suspension Buffer EZ and TELL-Seq™ Library Reagent Box 2 V1 adds an optional bead type, TELL Bead Plex.

Suspension Buffer EZ has reduced viscosity and is easy to pipette accurately. It further improves library yield. Suspension Buffer EZ requires larger input volume than previous Suspension Buffer in a reaction.

TELL-Seq™ Library Reagent Box 2 V1 contains either standard TELL Bead or new TELL Bead Plex. TELL Bead Plex beads use modified linked read barcode (index 1) sequences and provide more balanced nucleotide signal at index 1 position 3, 6, 7, 12,13 and 16. While TELL-Seq™ libraries generated with standard TELL Bead produce excellent Index 1 sequencing quality on all Illumina systems, TELL-Seq™ libraries generated with TELL Bead Plex show excellent Index 1 sequencing quality on a variety of sequencing platforms including Illumina and non-Illumina systems. TELL-Read pipeline v1.1 or above is required to analyze sequencing data generated from TELL-Seq™ libraries prepared with TELL Bead Plex and works for sequencing data generated from standard TELL Bead as well.

Index 1 sequencing profile of TELL-Seq libraries prepared with different type of TELL Beads



2. Kit Contents

TELL-Seq™ Library Prep Kit, Standard Size (2 Boxes)

Box 1 of 2: TELL-Seq™ Library Reagent Box 1 V1 (PN 100035)

NOTE: Do not freeze and thaw Box 1 reagents for more than 6 times.

Component Name	Cap Color	Volume (μL)	Storage Temperature
5× Reaction Buffer	CAP Blue	120	-25°C to -15°C
Barcoding Enzyme	CAP Black	24	-25°C to -15°C
Cofactor II	CAP Amber	120	-25°C to -15°C
Exonuclease	CAP Yellow	12	-25°C to -15°C
Stabilizer	CAP Violet	12	-25°C to -15°C
Suspension Buffer EZ	CAP Natural	180	-25°C to -15°C
Tagging Enzyme	CAP Red	24	-25°C to -15°C
2× PCR Master Mix	CAP Pink	150	-25°C to -15°C
Enhancer	CAP Green	18	-25°C to -15°C
10× Primer I ^a	CAP White	30	-25°C to -15°C

^a For use with 10× Primer II in any TELL-Seq™ Library Multiplex Primer Kit together for library amplification.

Box 2 of 2: TELL-Seq™ Library Reagent Box 2 V1 (PN 100036)

Component Name	Cap Color	Volume (μL)	Storage Temperature
TELL Bead or TELL Bead Plex ^b	CAP Orange	76	2°C to 8°C
Wash Solution	CAP White	4500	2°C to 8°C
Stop Solution ^c	CAP Natural	960	2°C to 25°C

^b TELL Bead works well on Illumina Sequencing Systems; TELL Bead Plex works well on both Illumina and non-Illumina Sequencing Systems.

^c Prior to use, if the Stop Solution is not clear or has white precipitates, warm the tube up at 37°C. Vortex to dissolve any precipitate. After the first use, store resuspended Stop Solution at room temperature for future use.









PRO TIP: FOUR TELL-Seq™ Library Prep Kits, Standard Size including both Box 1 and Box 2 can pair with **ONE** of any TELL-Seq™ Library Multiplex Primer Kits.



CAUTION









TELL-Read pipeline v1.1 or above is required to analyze sequencing data generated from TELL-Seq™ libraries prepared with TELL Bead Plex and works for data generated from standard TELL Bead as well.

TELL-Seq™ Library Multiplex Primer (1-8) Kit (PN 100003)

Component Name	Cap Color	Volume (μL)	Storage Temperature
10× Primer II, T501	 Blue	15	-25°C to -15°C
10× Primer II, T502	 Black	15	-25°C to -15°C
10× Primer II, T503	 Green	15	-25°C to -15°C
10× Primer II, T504	 Yellow	15	-25°C to -15°C
10× Primer II, T505	 Violet	15	-25°C to -15°C
10× Primer II, T506	 Natural	15	-25°C to -15°C
10× Primer II, T507	 Red	15	-25°C to -15°C
10× Primer II, T508	 Orange	15	-25°C to -15°C






PRO TIP: ONE TELL-Seq™ Library Multiplex Primer (1-8) Kit contains enough primers to be used with **FOUR** TELL-Seq™ Library Prep Kits, Standard Size.

TELL-Seq™ Library Multiplex Primer (9-16) Kit (PN 100009)

Component Name	Cap Color	Volume (μL)	Storage Temperature
10× Primer II, T509	 Blue	15	-25°C to -15°C
10× Primer II, T510	 Amber	15	-25°C to -15°C
10× Primer II, T511	 Green	15	-25°C to -15°C
10× Primer II, T512	 Yellow	15	-25°C to -15°C
10× Primer II, T513	 Violet	15	-25°C to -15°C
10× Primer II, T514	 Orange	15	-25°C to -15°C
10× Primer II, T515	 Red	15	-25°C to -15°C
10× Primer II, T516	 Natural	15	-25°C to -15°C

PRO TIP: ONE TELL-Seq™ Library Multiplex Primer (9-16) Kit contains enough primers to be used with **FOUR** TELL-Seq™ Library Prep Kits, Standard Size.

TELL-Seq™ Library Multiplex Primer (17-24) Kit (PN 100010)

Component Name	Cap Color	Volume (μL)	Storage Temperature
10× Primer II, T517	 Amber	15	-25°C to -15°C
10× Primer II, T518	 Blue	15	-25°C to -15°C
10× Primer II, T519	 Yellow	15	-25°C to -15°C
10× Primer II, T520	 Green	15	-25°C to -15°C
10× Primer II, T521	 Black	15	-25°C to -15°C
10× Primer II, T522	 Violet	15	-25°C to -15°C
10× Primer II, T523	 Orange	15	-25°C to -15°C
10× Primer II, T524	 Red	15	-25°C to -15°C

PRO TIP: ONE TELL-Seq™ Library Multiplex Primer (17-24) Kit contains enough primers to be used with **FOUR** TELL-Seq™ Library Prep Kits, Standard Size.





TELL-Seq™ Library Multiplex Primer C-series (1-96) Plate, (PN 100043)

Contains 10× Primer II, C501 to C596 on a 96-well plate. All primers are at 10× ready-to-use concentration and 30µL per well.

	1	2	3	4	5	6	7	8	9	10	11	12
A	C501	C509	C517	C525	C533	C541	C549	C557	C565	C573	C581	C589
B	C502	C510	C518	C526	C534	C542	C550	C558	C566	C574	C582	C590
C	C503	C511	C519	C527	C535	C543	C551	C559	C567	C575	C583	C591
D	C504	C512	C520	C528	C536	C544	C552	C560	C568	C576	C584	C592
E	C505	C513	C521	C529	C537	C545	C553	C561	C569	C577	C585	C593
F	C506	C514	C522	C530	C538	C546	C554	C562	C570	C578	C586	C594
G	C507	C515	C523	C531	C539	C547	C555	C563	C571	C579	C587	C595
H	C508	C516	C524	C532	C540	C548	C556	C564	C572	C580	C588	C596

NOTE: TELL-Seq™ Library Multiplex Primer C-series (1-96) Plate uses 10-base index sequences that differ from the TELL-Seq™ Library Multiplex Primer (1-8, 9-16, 17-24) kit (T-series) which uses 8-base index sequences.

TELL-Seq™ Illumina® Sequencing Primer Kit (PN 100004)

Component Name	Cap Color	Concentration	Volume (µL)	Storage Temperature
Read 1 Primer	 Black	100µM	50	-25°C to -15°C
Read 2 Primer	 White	100µM	50	-25°C to -15°C
Index 1 Primer	 Red	100µM	50	-25°C to -15°C
Index 2 Primer	 Yellow	100µM	50	-25°C to -15°C











PRO TIP: The minimum number of sequencing runs that can be performed using the amount of sequencing primers provided vary based on the sequencing system (see below).

Sequencing System	Number of runs	Is custom Index 2 Primer required?
NovaSeq	4	v1 reagent: No ; v1.5 reagent: Yes
HiSeq 3000/4000	2	Yes
HiSeq 2000/2500	5	No
NextSeq	8	Yes
MiSeq	16	No
MiniSeq	8	Yes

TELL-Seq™ Library Prep Kit, HT24 (2 Boxes)




Box 1 of 2: TELL-Seq™ Library Reagent Box 1 V1, HT24 (PN 100037)

NOTE: Do not freeze and thaw Box 1 reagents for more than 6 times.

Component Name	Cap Color	Volume (μL)	Storage Temperature
5× Reaction Buffer	 Blue	720	-25°C to -15°C
Barcoding Enzyme	 Black	144	-25°C to -15°C
Cofactor II	 Amber	720	-25°C to -15°C
Exonuclease	 Yellow	72	-25°C to -15°C
Stabilizer	 Violet	72	-25°C to -15°C
Suspension Buffer EZ	 Natural	1080	-25°C to -15°C
Tagging Enzyme	 Red	144	-25°C to -15°C
2× PCR Master Mix	 Pink	900	-25°C to -15°C
Enhancer	 Green	108	-25°C to -15°C
10× Primer I ^a	 White	180	-25°C to -15°C

^a For use with 10× Primer II in the TELL-Seq™ Library Multiplex Primer Kit together for library amplification.

Box 2 of 2: TELL-Seq™ Library Reagent Box 2 V1, HT24 (PN 100038)

Component Name	Cap Color	Volume	Storage Temperature
TELL Bead or TELL Bead Plex ^b	 Orange	456 μL	2°C to 8°C
Wash Solution	 Blue	28.5 mL	2°C to 8°C
Stop Solution ^c	 White	5.76 mL	2°C to 25°C

^b TELL Bead works well on Illumina Sequencing Systems; TELL Bead Plex works well on both Illumina and non-Illumina Sequencing Systems.

^c Prior to use, if the Stop Solution is not clear or has white precipitates, warm the tube up at 37°C. Vortex to dissolve any precipitate. After the first use, store resuspended Stop Solution at room temperature for future use.





PRO TIP: TWO TELL-Seq™ Library Prep Kits, HT24 including both Box 1 and Box 2 can pair with **THREE** of any TELL-Seq™ Library Multiplex Primer Kits.



CAUTION

TELL-Read pipeline v1.1 or above is required to analyze sequencing data generated from TELL-Seq™ libraries prepared with TELL Bead Plex and works for data generated from standard TELL Bead as well.

TELL-Seq™ Illumina® Sequencing Primer Kit, HT (PN 100013)

Component Name	Cap Color	Concentration	Volume (μL)	Storage Temperature
Read 1 Primer	 Black	100μM	300	-25°C to -15°C
Read 2 Primer	 White	100μM	300	-25°C to -15°C
Index 1 Primer	 Red	100μM	300	-25°C to -15°C
Index 2 Primer	 Yellow	100μM	300	-25°C to -15°C

PRO TIP: The minimum number of sequencing runs that can be performed using the amount of sequencing primers provided vary based on the sequencing system (see below).

Sequencing System	Number of runs	Is custom Index 2 Primer required?
NovaSeq	24	v1 reagent: No ; v1.5 reagent: Yes
HiSeq 3000/4000	12	Yes
HiSeq 2000/2500	30	No
NextSeq	48	Yes
MiSeq	96	No
MiniSeq	48	Yes

3. Consumables and Equipment (not provided)

Consumables

Consumable	Supplier
0.2 mL PCR tube or strip tube, RNase and DNase free	General lab supplier
20 µL pipette tip (standard and wide orifice)	General lab supplier
200 µL pipette tip (standard and wide orifice)	General lab supplier
Ethanol 200 proof (absolute) for molecular biology (500 mL)	Sigma-Aldrich, # E7023
Nuclease-free water	General lab supplier
AMPure XP	Beckman, # A63880
Agilent Bioanalyzer High Sensitivity DNA Analysis Kit*	Agilent, # 5067-4626
TapeStation High Sensitivity D5000 ScreenTape Assay*	Agilent, # 5067-5592, #5067-5593
Qubit dsDNA HS Assay Kit	Thermo Fisher Scientific, # Q32851 or Q32854
Qubit Assay Tubes	Thermo Fisher Scientific, # Q32856
TE buffer, pH 8.0	General lab supplier


*Depends on which system is available in the user facility.

Equipment

Equipment	Supplier
Thermo Cycler	Applied Biosystems
Magnetic stand for 0.2 mL PCR tubes	General lab supplier
Tube Rotator	General lab supplier
Incubator (for 35°C)	General lab supplier
Vortexer	General lab supplier
Microcentrifuge	General lab supplier
Agilent Bioanalyzer*	Agilent
Agilent TapeStation*	Agilent
Qubit® Fluorometer 3.0 or higher	Thermo Fisher Scientific, # Q33216 or Q33218
Ice Bucket	General lab supplier

*Depends on which system is available in the user facility.

4. TELL-Seq™ Library Prep Workflow



5. Protocol

TELL-Seq™ Library Prep Kits are designed to generate up to 12 TELL-Seq™ WGS libraries using Standard Size kit and up to 72 TELL-Seq™ WGS libraries using HT24 kit (see table below). The following protocol describes library preparation procedures based on specified sample genome sizes. All other unspecified conditions will apply to all genome sizes.



CAUTION

TELL-Read pipeline v1.1 or above is required to analyze sequencing data generated from TELL-Seq™ libraries prepared with TELL Bead Plex and works for data generated from standard TELL Bead as well.

Barcode DNA

I. Samples

- Input genomic DNA (User)

Genome Size	Input Amount	Reaction Vol (μL)	Preps/ Standard Size Kit	Preps/ HT24 Kit
1 Mb – 50 Mb	0.5 ng	50	12	72
50 Mb – 100 Mb	1 ng	50	12	72
100 Mb – 200 Mb	1.5 ng	50	12	72
200 Mb – 1 Gb	2 – 3 ng	100	6	36
1 Gb – 5 Gb	3 – 5 ng	150	4	24

PRO TIP: Use 4-5 ng of input DNA for a human genome.

NOTE:

1. Genomic DNA should be stored and diluted in a Tris buffer with pH ranging from 7.5 to 8.0 or a low TE buffer (10mM Tris-HCl, 0.1 mM EDTA, pH 8.0).
2. The recommended amount of input genomic DNA is based on an average fragment size of ≥ 30 Kb. If the average size of genomic DNA is ~ 15 Kb only, reducing the amount of input per reaction may be necessary.

II. Preparation

1. Prepare the following consumables:


Item	Location	Storage	Instruction
5× Reaction Buffer CAP	Box 1	-25°C to -15°C	Thaw at room temperature. Flick the tube with a finger 4 to 5 times to mix, then centrifuge briefly. Keep on ice until needed.
Cofactor II CAP	Box 1	-25°C to -15°C	Vortex to mix, then centrifuge briefly. Keep at room temperature in the dark. Close the tube cap tightly after each use.
Barcoding Enzyme CAP	Box 1	-25°C to -15°C	Centrifuge briefly. Keep on ice until needed.
TELL Bead or TELL Bead Plex CAP	Box 2	2°C to 8°C	Centrifuge briefly. Keep on ice until needed. Close the tube cap tightly after each use to avoid any evaporation.
Suspension Buffer EZ CAP	Box 1	-25°C to -15°C	Thaw at room temperature. Vortex to mix, then centrifuge briefly. Keep at room temperature until needed.
Nuclease-free water	User		Keep at room temperature.
0.2 mL PCR tube or strip tube	User		Room temperature.
20 µL wide orifice pipette tips	User		Room temperature.
200 µL wide orifice pipette tips	User		Room temperature.

2. Set up a tube rotator in a 35°C incubator (see Step 7 of the Procedure Section).




CAUTION

Use wide orifice pipette tips to transfer and mix high molecular weight genomic DNA to avoid shearing the DNA. If wide orifice pipette tips are not available, cut 4mm-5mm off a standard pipette tip top with a sterile razor blade or a pair of sharp scissors before use.



A. 20µL pipette tips (uncut and cut)



B. 200µL pipette tips (uncut and cut)

III. Procedure


1. Vortex TELL Bead or TELL Bead Plex vigorously for at least 30 seconds. Pulse spin (centrifuge for no more than 1 second) to bring down the bead solution present on the lid or sides of the tube. Right before use, pipet the TELL Bead or TELL Bead Plex with a 200 μL tip up and down 5 times to make sure all the beads are resuspended properly.
2. In a 0.2 mL PCR tube, assemble each reaction in the following order.

Reagent	Volume per reaction (μL)		
	Small Genome (50 μL)	Medium Genome (100 μL)	Large Genome (150 μL)
5 \times Reaction Buffer CAP	10	20	30
Nuclease-free water	7 – X (X is the DNA vol)	14 – Y (Y is the DNA vol)	20 – Z (Z is the DNA vol)
Cofactor II CAP	10	20	30
TELL Bead or TELL Bead Plex CAP (~0.5M barcodes/ μL)	6	12	19


3. Mix well by pipetting up and down for 10 times or vortexing vigorously for 5 seconds and pulse spin to bring the solution down to the bottom. Add appropriate amount of Barcoding Enzyme.

Reagent	Volume per reaction (μL)		
	Small Genome	Medium Genome	Large Genome
Barcoding Enzyme CAP	2	4	6

4. Mix well by pipetting up and down for 8 times. Avoid introducing air bubbles when pipetting by keeping the pipette tip at the bottom of the solution in the tube.



A. Bubble issue



B. Properly mixed

- Use a wide orifice pipette tip, add following reagents to the sample tube.

Reagent	Volume per reaction (μL)		
	Small Genome	Medium Genome	Large Genome
Sample genomic DNA	X μL ($\leq 5 \mu\text{L}$)	Y μL ($\leq 10 \mu\text{L}$)	Z μL ($\leq 15 \mu\text{L}$)
Suspension Buffer EZ CAP	15	30	45

NOTE: Suspension Buffer EZ is highly viscous. Use caution and pipette slowly to ensure that correct volume is delivered.

- Set the pipette volume based on the genome size of the sample (small-35 μL , medium-70 μL , large-110 μL). Use a wide orifice pipette tip, gently mix the solution by **slowly** pipetting up and down 6-8 times. Avoid introducing many air bubbles when pipetting by keeping the pipette tip at the bottom of the solution in the tube.
- Place the sample tube on a tube rotator in a 35°C incubator and rotate slowly (10 -15 rpm) for 30 minutes.



Sample tubes placed on a Tube Rotator in a 35°C incubator.

Stabilize DNA

I. Preparation

- Prepare the following consumables:

Item	Location	Storage	Instruction
Stabilizer CAP	Box 1	-25°C to -15°C	Flick the tube 4 to 5 times to mix. Centrifuge briefly. Keep on ice until needed.

II. Procedure

1. Retrieve the sample tube from the 35°C incubator after incubation.
2. Add appropriate amount of Stabilizer into the tube.

Reagent	Volume per reaction (μL)		
	Small Genome	Medium Genome	Large Genome
Stabilizer CAP	1	2	3

3. Set pipette volume based on the genome size of the sample (small-35 μL, medium-70 μL, large-110 μL). Use a wide orifice pipette tip, gently mix the solution by **slowly** pipetting up and down 6-8 times. Avoid creating many bubbles.
4. Place the sample tube back on the tube rotator in the 35°C incubator and rotate it slowly (10 - 15 rpm) for 30 minutes.

Tagment DNA

I. Preparation

1. Prepare the following consumables:

Item	Location	Storage	Instruction
Tagging Enzyme CAP	Box 1	-25°C to -15°C	Flick the tube 4 to 5 times to mix. Centrifuge briefly. Keep on ice until needed.
Exonuclease CAP	Box 1	-25°C to -15°C	Flick the tube 4 to 5 times to mix. Centrifuge briefly. Keep on ice until needed.

II. Procedure

1. Retrieve the sample tube from the 35°C incubator after incubation.
2. Add Tagging Enzyme and Exonuclease into the tube.

Reagent	Volume per reaction (μL)		
	Small Genome	Medium Genome	Large Genome
Tagging Enzyme CAP	1	2	3
Exonuclease CAP	1	2	3

3. Set the pipette based on the genome size of the sample (small-35 μL, medium-70 μL, large-110 μL). Use a wide orifice pipette tip, gently mix the solution by **slowly** pipetting up and down for 8 to 10 times. For this step, the mixing needs to be very thorough. Avoid creating many bubbles.

- Place the sample tube back on the tube rotator in the 35°C incubator and rotate it slowly for 10 minutes. When necessary, different amounts of Tagging Enzyme can be used to adjust the library size.

NOTE: If a longer insert library is preferred, less amount of Tagging Enzyme can be used in the reaction. On the other hand, if a shorter insert library is preferred, twice as much as of Tagging Enzymes can be used in the reaction.

- Proceed to next step immediately after the incubation.

Wash Beads

I. Preparation

- Prepare the following consumables:

Item	Location	Storage	Instruction
Stop Solution			
CAP Natural in the standard size kit	Box 2	2°C to 25°C	Check for any precipitates. If present, incubate the buffer at 37°C for 10 minutes, and vortex until they dissolve. Store at room temperature for future use.
CAP White in the HT24 kit			
Wash Solution			
CAP White in standard size kit	Box 2	2°C to 8°C	Bring to room temperature.
CAP Blue in the HT24 kit			
0.2 mL PCR tube or strip tube	User		Room temperature.

- Set up a thermo cycler with the following program:
 - Set Preheat lid option to 100°C
 - 63°C forever

II. Procedure

- Place the sample tube on a magnetic stand for 1 minute or until the solution is clear.
- While the tube is on the magnetic stand, aspirate and discard the supernatant without disturbing beads.
- Remove the tube from the magnetic stand. Add 120 µL Wash Solution to the sample tube. Pipet to resuspend the beads. If necessary, pulse spin to bring the solution down.
- Place the sample tube back on the magnetic stand for 1 minute or until the solution is clear.
- While the tube is on the magnetic stand, aspirate and discard the supernatant without disturbing beads.
- Remove the tube from the magnetic stand. Add 80 µL of Stop Solution to the tube.

7. Pipet several times to resuspend the beads. If necessary, pulse spin to bring the solution down.
8. Incubate the tube at room temperature for 5 minutes.
9. Place the sample tube back on the magnetic stand for 1 minute or until the solution is clear.
10. While the tube is on the magnetic stand, aspirate and discard the supernatant without disturbing beads.
11. Remove the tube from the magnetic stand. Add 120 μ L Wash Solution to the PCR tube. Pipet to resuspend the beads.
12. **Transfer all the bead solution into a new 0.2ml PCR tube.**
13. Incubate the tube at 63°C on the PCR thermocycler for 3 minutes.
14. Place the new sample tube on the magnetic stand at room temperature for 1 minute or until the solution is clear.
15. While the tube is on the magnetic stand, aspirate and discard the supernatant without disturbing beads.
16. Remove the tube from the magnetic stand. Add 120 μ L Wash Solution to the PCR tube. Pipet to resuspend the beads. If necessary, pulse spin to bring the solution down.
17. Incubate the tube at 63°C on the PCR thermocycler for 3 minutes.
18. Place the sample tube on the magnetic stand at room temperature for 1 minute or until the solution is clear.
19. While the tube is on the magnetic stand, aspirate and discard the supernatant without disturbing beads. Use a P20 pipette to remove any remaining supernatant.
20. Remove the tube from the magnetic stand. Resuspend the beads in 20 μ L of Wash Solution.

NOTE:

This is a **SAFE STOPPING POINT**. The washed beads can be stored at 2°C to 8°C for two weeks.

Amplify Library

I. Preparation

1. Prepare the following consumables:

Item	Location	Storage	Instruction
2× PCR Master Mix CAP	Box 1	-25°C to -15°C	Thaw at room temperature. Flick the tube 4 to 5 times to mix, then centrifuge briefly. Keep on ice.
10× Primer I CAP	Box 1	-25°C to -15°C	Thaw at room temperature. Vortex to mix, then centrifuge briefly. Keep on ice.
10× Primer II, T5## (T-series)	Multiplex Primer Kit	-25°C to -15°C	Thaw at room temperature. Vortex to mix, then centrifuge briefly. Keep on ice.
Enhancer CAP	Box 1	-25°C to -15°C	Thaw at room temperature. Vortex to mix, then centrifuge briefly. Keep at room temperature.
Nuclease-free water	User	Room Temperature	Keep at room temperature.
0.2 mL PCR tube or strip tube	User		Room temperature.



CAUTION

For high throughput library preparation using Multiplex Primer C-series (1-96) Plate, please refer to the Appendix section, Usage of TELL-Seq™ Multiplex Primer Plate, for general guidance on plate handling to avoid cross contamination of index primers.

2. Set up Library Amplification Program (**LAP**) on a thermo cycler as follows:

- 63°C 2 minutes
- 72°C 2 minutes
- 98°C 30 seconds
- [98°C 15 seconds, 63°C 20 seconds, 72°C 30 seconds] x Cycle Number
- 72°C 3 minutes
- 4°C forever

NOTE:

The deeper sequencing depth for each TELL Bead, will lead to a higher linked read density and better performance. Hence, for a fixed number of sequencing reads, the fewer TELL Beads used for library amplification, the deeper sequencing depth per bead, which will lead to a better linked read result. However, if too few TELL Beads were used for library amplification, the library complexity would be low, and sequencing read duplication level would be high.

For GC-rich genomes (GC>60%), amplify one more cycle than would be done for samples with low GC content.

Genome Size	Vol of Beads Used (B) for PCR	PCR Volume	Cycle Number
1 Mb – 50 Mb	1 – 12 μ L	25 μ L	14 – 11
50 Mb – 100 Mb	6 – 10 μ L	25 μ L	12 – 10
100 Mb – 200 Mb	10 – 20 μ L	25 μ L	12 – 10
200 Mb – 1 Gb	15 – 20 μ L	50 μ L	11 – 9
1 Gb – 5 Gb	12 – 20 μ L	75 μ L	10 – 9

PRO TIP:

a) For *E. coli* (4.6Mb) use 2-3 μ L of TELL Beads and 13 cycles.

b) For human (3Gb) use 20 μ L of TELL Beads and 9 cycles.

II. Procedure

1. Vortex beads vigorously for 10 seconds to resuspend the beads. Pulse spin to bring solution down. Using a 20 μ L pipette tip, pipet the beads up and down 5 times to make sure all the beads are resuspended properly prior use. Immediately transfer the appropriate amount of bead solution (B in table above) to a new 0.2ml PCR tube.
2. If $B \leq 2 \mu\text{L}$, proceed to Step 5 directly.
3. If $B > 2 \mu\text{L}$, place the PCR tube on a magnetic stand for 1 minute or until the solution is clear.
4. While the tube is on the magnetic stand, remove and discard (B-2) μ L supernatant without disturbing beads. Remove the PCR tube from the magnet.
5. Add following reagents to the PCR tube containing the beads based on sample genome size.

Reagent	Volume per reaction (μ L)		
	Small Genome (25 μ L)	Medium Genome (50 μ L)	Large Genome (75 μ L)
Nuclease-free water	4 μ L	10 μ L	16 μ L
2 \times PCR Master Mix CAP	12.5 μ L	25 μ L	37.5 μ L
10 \times Primer I CAP	2.5 μ L	5 μ L	7.5 μ L
10 \times Primer II, T5##	2.5 μ L	5 μ L	7.5 μ L
Enhancer CAP	1.5 μ L	3 μ L	4.5 μ L

6. Mix well by vortexing or pipetting. Pulse spin to bring solution down.
7. Place the tube on the thermal cycler and run the **LAP** program (see above) with the appropriate number of cycles based on the genome size of the sample.
8. After PCR amplification, use 2 μ L PCR product for quality check on a Bioanalyzer or a TapeStation. See Qualify and Quantify Library section for instruction.

PRO TIP: If QC check shows the library yield is relatively low, put the tube with remaining PCR product back to the thermocycler and amplify for another one or two extra cycles before moving to Clean Up Library section.

NOTE:

This is a **SAFE STOPPING POINT**. The PCR product can be stored at -25°C to -15°C for one month.

Clean Up Library

I. Preparation

1. Prepare the following consumables:

Item	Location	Storage	Instruction
Ethanol 200 proof (absolute) for molecular biology	User	Room Temperature	
Fresh 75% (v/v) ethanol	User	Room Temperature	Require 400 μ L per sample. Mix 1.5 mL Ethanol (200 proof) with 0.5 mL Nuclease-free water. Vortex to mix and keep at room temperature.
AMPure XP	User	2°C to 8°C	Bring it to room temperature for at least 20 minutes and vortex vigorously to resuspend the beads before use.
Nuclease-free water	User	Room Temperature	Keep at room temperature.
TE buffer, pH 8.0	User	Room Temperature	Keep at room temperature.
0.2 mL PCR tube or strip tube	User		Room Temperature

NOTE:

There are two different cleanup procedures based on the library size profile.

1. TELL-Seq™ library sizes are broader than standard Illumina libraries but most under 1kb usually. Single SPRI cleanup in the Step II is good enough to remove adapter dimers in the library.
2. For large genome samples, their library size can be much wider sometimes. When there are greater than 40% library molecules are over 1 kb, a double SPRI cleanup in the Step III is recommended to remove both adapter dimers and >1kb library products.

II. Standard Procedure (Single SPRI Cleanup)

1. Briefly centrifuge the sample PCR tube to bring all solution down.
2. Place the tube on the magnetic stand for 1 minute or until the solution is clear.
3. While the tube is on the magnetic stand, transfer the supernatant to a new 0.2 mL PCR tube without disturbing beads.
4. Measure the volume of transferred supernatant (PCR product) with a pipette.
5. Add following reagents into the PCR product to a total volume of 100 µL.

Reagent	Volume per reaction
PCR product	Measured Volume
Nuclease-free water	To final 100 µL total

6. Vortex vigorously to resuspend the AMPure XP solution and add 78 µL AMPure XP into the 100 µL PCR product.
7. Mix by pipetting up and down 10 times.
8. Incubate at room temperature for 5 minutes.
9. Place the tube on the magnetic stand for 1 minute or until the solution is clear.
10. Use a pipette, aspirate and discard the supernatant without disturbing AMPure beads.
11. While keeping the tube on the magnetic stand, add 200 µL of freshly prepared 75% ethanol into the tube. Let it sit for 30 seconds.
12. Aspirate and discard the supernatant without disturbing beads.
13. Repeat steps 11-12 one more time, keeping the tube on the magnetic stand for the whole time. Use a P20 pipette to remove any remaining supernatant.
14. Keep the tube on the magnetic stand with the cap open and allow the tube to dry for 1-2 minutes to evaporate traces of ethanol. DON'T over-dry the beads.
15. Remove the tube from the magnetic stand and add 25 µL TE buffer to the beads.
16. Pipette or vortex to resuspend the beads. Incubate at room temperature for 5 minutes.
17. Place the tube on the magnetic stand for 1 minute or until the solution is clear.
18. With the tube still on the magnetic stand, carefully transfer 23 µL of the supernatant to a new tube. Be careful not to disturb the beads.
19. The supernatant contains the TELL-Seq™ library. Proceed to **Qualify and Quantify Library**.

NOTE:

This is a **SAFE STOPPING POINT**. The purified TELL-Seq™ library can be stored at -25°C to -15°C for six months.

III. Double SPRI Cleanup Procedure (for library with >40% molecules over 1kb size)

1. Briefly centrifuge the sample PCR tube to bring all solution down.
2. Place the PCR tube on the magnetic stand for 1 minute or until the solution is clear.
3. While the tube is on the magnetic stand, transfer the supernatant to a new 0.2 mL PCR tube without disturbing beads.
4. Measure the volume of transferred supernatant (PCR product) with a pipette.
5. Add the following reagents into the PCR product to a total volume of 100 μ L.

Reagent	Volume per reaction
PCR product	Measured Volume
Nuclease-free water	To final 100 μ L total

6. Vortex vigorously to resuspend the AMPure XP solution and add 50 μ L AMPure XP into the 100 μ L PCR product.
7. Mix well by pipetting up and down 10 times.
8. Incubate at room temperature for 5 minutes.
9. Place the tube on the magnetic stand for 1 minute or until the solution is clear.
10. Transfer the supernatant to a new 0.2 mL PCR tube without disturbing AMPure beads.
11. Add 28 μ L AMPure XP into the supernatant transferred to the new PCR tube.
12. Mix well by pipetting up and down 10 times.
13. Incubate at room temperature for 5 minutes.
14. Place the tube on the magnetic stand for 1 minute or until the solution is clear.
15. Aspirate and discard the supernatant without disturbing AMPure beads.
16. While keeping the tube on the magnetic stand, add 200 μ L of freshly prepared 75% ethanol into the tube. Let it sit for 30 seconds.
17. Aspirate and discard the supernatant without disturbing beads.
18. Repeat steps 16-17 one more time, keeping the tube on the magnetic stand for the whole time.
19. Keep the tube on the magnetic stand with the cap open and allow the tube to dry for 1-2 minutes to evaporate traces of ethanol. DON'T over-dry the beads.
20. Remove the tube from the magnetic stand and add 25 μ L TE buffer to the beads.
21. Pipette or vortex to resuspend the beads. Incubate at room temperature for 5 minutes.
22. Place the tube on the magnetic stand for 1 minute or until the solution is clear.
23. With the tube still on the magnetic stand, carefully transfer 23 μ L of the supernatant to a new tube. Be careful not to disturb the beads.
24. The supernatant contains the TELL-Seq™ library. Proceed to **Qualify and Quantify Library**.

NOTE:

This is a **SAFE STOPPING POINT**. The purified TELL-Seq™ library can be stored at -25°C to -15°C for six months.

Qualify and Quantify Library

I. Consumables

- Agilent High Sensitivity DNA Kit **or** TapeStation High Sensitivity D5000 ScreenTape Assay (User)
- Qubit dsDNA HS Assay Kit (User)
- Qubit assay tube (User)
- TE buffer, pH 8.0 (User)

NOTE:

Standard qPCR library quantitation assay for Illumina system works for TELL-Seq™ library, but it is not required.

II. Preparation

1. Prepare the necessary consumables as required by Bioanalyzer or TapeStation and Qubit.

III. Procedure

1. Use 1 μ L of library for Agilent High Sensitivity DNA Kit or 2 μ L of library for TapeStation High Sensitivity D5000 ScreenTape Assay.
2. Check the saved uncleaned PCR product from the Amplify Library section at the same time. Uncleaned PCR product may have a high level of primer dimer and adapter dimer. It requires a two-fold dilution with nuclease-free water before loading onto a Bioanalyzer chip or TapeStation tape to avoid interfering with lower marker signal.
3. To determine the library concentration, set the Region on the Bioanalyzer or TapeStation analysis software from 150 bp to 1000 bp. Record sample Concentration (nM) for this region (see Figure 1). To determine the library size, set the Region from 150 bp to 3000 bp. Record sample Average Size (bp) as Library Size. A good-sized library should have most library fragments under 1000 bp.



CAUTION

The concentration reading from the Bioanalyzer (or TapeStation) should be used as a starting point to make necessary dilution or library pooling for sequencing. Verify the concentration of the final diluted sequencing library or library pool with a Qubit dsDNA HS Assay kit (see Step 6).




Figure 1. An example of cleaned up library profile from a TapeStation High Sensitivity D5000 ScreenTape assay.

- Library can be sequenced immediately or stored at -25°C to -15°C.

NOTE:

Occasionally, there is a detectable residual level of adapter dimer in a cleaned-up library (see Figure 2). An additional round of Single SPRI Cleanup as described in Clean Up Library section is recommended in this case.




Figure 2. A library with detectable residual adapter dimer (arrow) after cleanup (TapeStation High Sensitivity D5000 ScreenTape assay).

5. When sequencing, dilute the library using TE buffer to the concentration recommended by each Illumina® sequencing system. Make a diluted library pool for sequencing if more than one library will be sequenced in the same run.
6. Use 4 µL **diluted** sequencing library or library pool to check the concentration with the Qubit dsDNA HS Assay Kit. Use the Library Size value measured from the Bioanalyzer (or TapeStation) for conversion of mass concentration into molar concentration.

A = Mass Concentration (ng/µL)

S = Library Size (bp)

$$\text{Molar Concentration (nM)} = (A * 1,000,000) / (S * 650)$$

7. Adjust the volume needed in the sequencing preparation if the library concentration measured by Qubit is different from the recommended concentration by more than 10%.

6. Appendix

Ultralow Input TELL-Seq™ Library Preparation

In a standard input TELL-Seq™ library procedure above, there will be 3 to 8 high molecular weight DNA fragments captured by one TELL Bead on average. When available sample DNA is rare or some cases, such as, mixed samples or targeted sequencing, prefer one HMW DNA fragment input to one TELL Bead (i.e. one unique barcode), the standard library protocol can be adjusted with lower input DNA and/or more TELL Beads to decrease the HMW DNA to TELL Bead ratio. Below are modified conditions for ultralow input TELL-Seq™ library preparation for small genome. All the other steps should follow the standard input library preparation procedure for small genome without changes.

- DNA input for small genome


Genome Size	Input Amount	Reaction Vol (μL)
1 Mb – 15 Mb	0.1 ng	50
16 Mb – 30 Mb	0.2 ng	50

- Volume of TELL Beads used in library amplification for small genome

Genome Size	Vol of Beads Used (B) for PCR	PCR Volume	Cycle Number
1 Mb – 15 Mb	5 - 20 μL	25 μL	14
16 Mb – 30 Mb	10 – 20 μL	25 μL	13


PRO TIP: For *E. coli* (4.6Mb) use 7.5 μL of TELL Beads and 14 cycles.

TELL-Seq™ Library Structure and Sequencing Scheme



Index 1 contains 18-base TELL Bead sequences, which must be sequenced completely. Index 2 contains 8-base sample index primer sequences used in library amplification. Paired end sequencing is preferred. Minimal read length requirement is 2x96; Maximum read length requirement is 2x150.

Example of Illumina® Sequencing % Base by Cycle Chart with TELL Bead & T-series Index II



Illumina® Sequencing Guide



CAUTION

TELL-Read pipeline v1.1 or above is required to analyze sequencing data generated from libraries prepared with TELL Bead Plex and works for data generated from standard TELL Bead as well.

1. Dilute TELL-Seq™ library according to Illumina® sequencing platform specific concentration and volume.
2. Libraries may be pooled together for sequencing when different multiplex primers are used in the library amplification step.
3. Custom sequencing primers are required to sequence TELL-Seq™ libraries and provided in the TELL-Seq™ Illumina Sequencing Primer Kit.

TELL-Seq™ Illumina Sequencing Primer Kits

Component Name	Concentration	Storage Temperature
Read 1 Primer	100 µM	-25°C to -15°C
Read 2 Primer	100 µM	-25°C to -15°C
Index 1 Primer	100 µM	-25°C to -15°C
Index 2 Primer	100 µM	-25°C to -15°C

4. These custom sequencing primers can be loaded into the specified wells for custom primers. Alternatively, they can also be loaded into corresponding standard Illumina® sequencing primer wells when an Illumina® PhiX control library is spiked in a sequencing run.
5. Custom Index 2 primer is only needed when multiple TELL-Seq™ libraries with different multiplex primers are pooled for sequencing and when a sequencer requires an i5 index sequencing primer. **For MiSeq, HiSeq 2000/2500 and NovaSeq v1 reagents, custom Index 2 Primer is not required.**
6. The minimum number of sequencing runs can be performed using the amount of sequencing primers provided are varied based on the sequencing system.

Sequencing System	Is custom Index 2 Primer required?
NovaSeq	v1 reagent: No ; v1.5 reagent: Yes
HiSeq 3000/4000	Yes
HiSeq 2000/2500	No
NextSeq	Yes
MiSeq	No
MiniSeq	Yes

Illumina® Sequencing Read Length Recommendation

1. Paired end sequencing is recommended.
2. TELL-Seq™ library Index 1 is 18-base, Index 2 (T-series) is 8-base. There are total 26-base for both indexes compared to total 16-base for standard Illumina dual index. The extra 10-cycle required for sequencing TELL-Seq™ library index need to be deducted from read 1 and read 2 sequencing cycles evenly. Since Illumina sequencing reagent guarantee 2 extra cycles (except NovaSeq v1.5 reagents), 4-cycle for read 1 and 4-cycle for read 2 need to be deducted, respectively. Recommended sequencing length is 2×96 PE with a 200-cycle kit or 2×146 PE with a 300-cycle kit for dual index run; 2x100 PE or 2x150 PE for a single sample run without need for Index 2 read.
3. For NovaSeq v1.5 reagents, there are extra reagents provided for index sequencing. Recommended sequencing length is 2×100 PE with a 200-cycle kit or 2×150 PE with a 300-cycle kit for dual index run.

Sequencing Depth Consideration

Adequate sequencing depth is required to get enough TELL Bead coverage. The more TELL Beads used in library amplification to generate a TELL-Seq™ library, the more sequencing reads will be required to get the desired sequencing depth. However, the fewer TELL Beads used for library amplification, the lower the library complexity will be, which may lead to a higher duplication rate of sequencing reads. The

balance between TELL Beads used and TELL-Seq™ library complexity required may depend on the genome size and application.

For *de novo* assembly application, at least 60× genome coverage of the sample is recommended in general. For scaffolding application, 30× genome coverage is recommended. For human whole genome phasing application, at least 500 million cluster reads per sample from a 2x146 PE or 2x150PE run are recommended.

Library Multiplex Primer Index Sequences (i.e. Index 2 Sequences): T-series (8-base)

Library Multiplex Primer	For Sample Sheet NovaSeq v1, MiSeq, HiSeq2000/2500	For Sample Sheet NovaSeq v1.5, Next Seq, MiniSeq, HiSeq3000/4000
T501	TGAACCTT	AAGGTTCA
T502	TGCTAAGT	ACTTAGCA
T503	TGTTCTCT	AGAGAACA
T504	TAAGACAC	GTGTCTTA
T505	CTAATCGA	TCGATTAG
T506	CTAGAACA	TGTTCTAG
T507	TAAGTTCC	GGAACTTA
T508	TAGACCTA	TAGGTCTA
T509	CATCCGAA	TTCGGATG
T510	TTATGAGT	ACTCATAA
T511	AGAGGCGC	GCGCCTCT
T512	TAGCCGCG	CGCGGCTA
T513	ACGAATAA	TTATTTCGT
T514	TTCGTAGG	CCTACGAA
T515	GATCTGCT	AGCAGATC
T516	CGCTCCGC	GCGGAGCG
T517	AGGCTATA	TATAGCCT
T518	GCCTCTAT	ATAGAGGC
T519	AGGATAGG	CCTATCCT
T520	TCAGAGCC	GGCTCTGA
T521	CTTCGCCT	AGGCGAAG
T522	TAAGATTA	TAATCTTA
T523	AGTAAGTA	TACTTACT
T524	GACTTCCT	AGGAAGTC

Library Multiplex Primer Index Sequences (i.e. Index 2 Sequences): C-series (10-base)

Library Multiplex Primer	For Sample Sheet NovaSeq v1, MiSeq, HiSeq2000/2500	For Sample Sheet NovaSeq v1.5, Next Seq, MiniSeq, HiSeq3000/4000
C501	ACGTACGTAC	GTACGTACGT
C502	CATGCATGCA	TGCATGCATG
C503	GTACGTACGT	ACGTACGTAC
C504	TGCATGCATG	CATGCATGCA
C505	ATGCTGATCA	TGATCAGCAT
C506	CACAGCTGTG	CACAGCTGTG
C507	GCTGATCAGC	GCTGATCAGC
C508	TGATCAGCAT	ATGCTGATCA
C509	ATTCAATACT	AGTATTGAAT
C510	CTAGCGCTAG	CTAGCGCTAG
C511	GCTAGTAGTA	TACTACTAGC
C512	TCCAATCAAG	CTTGATTGGA
C513	AATATTGCTG	CAGCAATATT
C514	CGTCGTTACG	CGTAACGACG
C515	GATTGATTCC	GGAATCAATC
C516	TCTAACAATG	CATTGTTAGA
C517	AGAATTGTCA	TGACAATTCT
C518	CTCAGCAATT	AATTGCTGAG
C519	GGTCCTTGTC	GACAAGGACC
C520	AGGCCTGACA	TGTCAGGCCT
C521	CTCCTAGTGG	CCACTAGGAG
C522	GGTTACAGCT	AGCTGTAACC
C523	CTGATTGGCG	CGCCAATCAG
C524	ATTGGTTAGA	TCTAACCAAT
C525	CCATTCAACT	AGTTGAATGG
C526	CAGTATTGAC	GTCAATACTG
C527	GAGTCCTCAA	TTGAGGACTC
C528	AGCTACTACT	AGTAGTAGCT
C529	TAGCTAGCGC	GCGCTAGCTA
C530	GATGCAACAC	GTGTTGCATC
C531	CCTCAGTACA	TGTACTGAGG
C532	CGGTAATTCA	TGAATTACCG

C533	CGCAATGGCT	AGCCATTGCG
C534	GTACGTTGAA	TTCAACGTAC
C535	TTGATCAGTA	TACTGATCAA
C536	GGCCTAACAA	TTGTTAGGCC
C537	GTTGTTGGAA	TTCCAACAAC
C538	TACGTTGGAC	GTCCAACGTA
C539	ACACCATGCA	TGCATGGTGT
C540	GCAATAGTAC	GTACTATTGC
C541	ACGCAGCCAG	CTGGCTGCGT
C542	CGAGTTGACG	CGTCAACTCG
C543	CGTGGCTGAA	TTCAGCCACG
C544	TCTCAAGGAC	GTCCTTGAGA
C545	CCTAGGCACT	AGTGCCTAGG
C546	CTGCGGTAAT	ATTACCGCAG
C547	GGCACTACCA	TGGTAGTGCC
C548	GCTCAATCAA	TTGATTGAGC
C549	AGGCACACAC	GTGTGTGCCT
C550	CCTGGCAAGA	TCTTGCCAGG
C551	TAATTGGTAG	CTACCAATTA
C552	GCCAACAAGT	ACTTGTTGGC
C553	ATGGCTTATA	TATAAGCCAT
C554	GCATGGCCTT	AAGGCCATGC
C555	ACAATACTGG	CCAGTATTGT
C556	GGATTGGACT	AGTCCAATCC
C557	ACTGTA CTAT	ATAGTACAGT
C558	CAGCTGTGAG	CTCACAGCTG
C559	CTTGAGGACC	GGTCCTCAAG
C560	GGTACAATAG	CTATTGTACC
C561	CTGACTACTA	TAGTAGTCAG
C562	TCAACCATGG	CCATGGTTGA
C563	ATTATAACCG	CGGTTATAAT
C564	ACTAGTCCTT	AAGGACTAGT
C565	ACTTGGACGT	ACGTCCAAGT
C566	ATGGTTAGGA	TCCTAACCAT
C567	ATGGTACCAA	TTGGTACCAT
C568	GAATTGACTC	GAGTCAATTC
C569	AGCAACCAGG	CCTGGTTGCT
C570	TACTGTGCTG	CAGCACAGTA
C571	CAACAACGTC	GACGTTGTTG

C572	CAGTAGCGCT	AGCGCTACTG
C573	ATTACCAATC	GATTGGTAAT
C574	TAAGGACCGC	GCGGTCCTTA
C575	ACACGTACCG	CGGTACGTGT
C576	CAACGTTGTT	AACAACGTTG
C577	ATTGTGCTGA	TCAGCACAAT
C578	GTACCAACAG	CTGTTGGTAC
C579	TTGTCAAGGA	TCCTTGACAA
C580	CTTGACGTA	TACGTACAAG
C581	TGCCTTGTA	TTACAAGGCA
C582	TAGTAGCTTA	TAAGCTACTA
C583	GACCGCAATG	CATTGCGGTC
C584	CTACTAGCTT	AAGCTAGTAG
C585	AGCACACGTT	AACGTGTGCT
C586	TGTTATAAGC	GCTTATAACA
C587	GTTGCCAAGT	ACTTGGCAAC
C588	CTGGCAACCG	CGGTTGCCAG
C589	TTAGGCCTTA	TAAGGCCTAA
C590	CGCAGCACAG	CTGTGCTGCG
C591	CTAGGCACAA	TTGTGCCTAG
C592	TGTTGTACAG	CTGTACAACA
C593	CTAACGTGGC	GCCACGTTAG
C594	GCGTACTGGT	ACCAGTACGC
C595	GGCCTGAATT	AATTCAGGCC
C596	CATGCTCGAG	CTCGAGCATG

Usage of TELL-Seq™ Multiplex Primer Plate

When TELL-Seq™ Library Multiplex Primer C-series (1-96) Plate is used for high throughput library preparation, extra precautions are required to avoid cross-contamination of any primers. All primers (Primer II) are at the 10× ready-to-use concentration. The amount of primer in each well are sufficient for multiple reactions. If desirable, single use aliquots can be prepared from the TELL-Seq™ Library Multiplex Primer plate to avoid cross contamination.



CAUTION

C-series multiplex primers have 10-base index sequences and require 10-cycle index 2 sequencing.

Following instructions are served as a general guidance. Detailed operation procedure is required to be established by individual high throughput operator before using the plate.

1. Retrieve the Multiplex Primer C-series 96-well plate from -20°C storage.
2. Ensure that the multiplex primers (C501 – C596) are thawed with no visible ice present. Centrifuge the plate at 1000 rpm for 1 minute.
3. Carefully remove the adhesive plate seal or film.
4. Using a new pipette tip(s), pipette the required amount (depending on the sample genome size) of 10x Primer II from each well on the 96-well multiplex primer plate for the number of samples being processed.
5. Refer to the table below for individual multiplex primer locations on the plate.

	1	2	3	4	5	6	7	8	9	10	11	12
A	C501	C509	C517	C525	C533	C541	C549	C557	C565	C573	C581	C589
B	C502	C510	C518	C526	C534	C542	C550	C558	C566	C574	C582	C590
C	C503	C511	C519	C527	C535	C543	C551	C559	C567	C575	C583	C591
D	C504	C512	C520	C528	C536	C544	C552	C560	C568	C576	C584	C592
E	C505	C513	C521	C529	C537	C545	C553	C561	C569	C577	C585	C593
F	C506	C514	C522	C530	C538	C546	C554	C562	C570	C578	C586	C594
G	C507	C515	C523	C531	C539	C547	C555	C563	C571	C579	C587	C595
H	C508	C516	C524	C532	C540	C548	C556	C564	C572	C580	C588	C596



CAUTION

When preparing multiple sample libraries, ensure that appropriate multiplex primer II is added to the correctly labeled well to avoid sample mix up.

6. If processing less than 96 samples, circle the wells that are used with a marker to help indicate and visualize where the samples are located.
7. Slowly pipet up and down 5 times to mix.
8. Seal the plate with a new adhesive plate seal and centrifuge the plate at 1000 rpm for 1 minute.